



Uploaded to the VFC Website

▶▶▶▶ 2021 ◀◀◀◀

This Document has been provided to you courtesy of Veterans-For-Change!

Feel free to pass to any veteran who might be able to use this information!

For thousands more files like this and hundreds of links to useful information, and hundreds of "Frequently Asked Questions, please go to:

[Veterans-For-Change](#)

If Veterans don't help Veterans, who will?

Note:

VFC is not liable for source information in this document, it is merely provided as a courtesy to our members & subscribers.



FORMULA see Table 1 MW: see Table 1 CAS: see Table 1 RTECS: see Table 1

METHOD: 2549, Issue 1		EVALUATION: PARTIAL		Issue 1: 15 May 1996	
OSHA: NIOSH: varies with compound ACGIH:		PROPERTIES: See Table 1			
SYNONYMS: VOCs; See individual compounds in Table 1					
SAMPLING			MEASUREMENT		
SAMPLER:	THERMAL DESORPTION TUBE (multi-bed sorbent tubes containing graphitized carbons and carbon molecular sieve sorbents [See Appendix])		TECHNIQUE:	THERMAL DESORPTION, GAS CHROMATOGRAPHY, MASS SPECTROMETRY	
FLOW RATE:	0.01 to 0.05 L/min		ANALYTE:	See Table 1	
VOL-MIN:	1 L		DESORPTION:	Thermal desorption	
-MAX:	6 L		INJECTION VOLUME:	Defined by desorption split flows (See Appendix)	
SHIPMENT:	Ambient in storage containers		TEMPERATURE-DESORPTION:	300 °C for 10 min.	
SAMPLE STABILITY:	Compound dependent (store @ -10 °C)		-DETECTOR (MS):	280 °C	
BLANKS:	1 to 3 per set		-COLUMN:	35 °C for 4 min; 8 °C/min to 150 °C, 15 °C/min to 300 °C	
ACCURACY			CARRIER GAS:	Helium	
RANGE STUDIED:	not applicable		COLUMN:	30 meter DB-1, 0.25-mm ID, 1.0- μ m film, or equivalent	
BIAS:	not applicable		CALIBRATION:	Identification based on mass spectra interpretation and computerized library searches.	
OVERALL PRECISION (\hat{S}_{rT}):	not applicable		RANGE:	not applicable	
ACCURACY:	not applicable		ESTIMATED LOD:	100 ng per tube or less	
			PRECISION (\hat{S}_r):	not applicable	
APPLICABILITY: This method has been used for the characterization of environments containing mixtures of volatile organic compounds (See Table 1). The sampling has been conducted using multi-bed thermal desorption tubes. The analysis procedure has been able to identify a wide range of organic compounds, based on operator expertise and library searching.					
INTERFERENCES: Compounds which coelute on the chromatographic column may present an interference in the identification of each compound. By appropriate use of background subtraction, the mass spectrometrist may be able to obtain more representative spectra of each compound and provide a tentative identity (See Table 1).					
OTHER METHODS: Other methods have been published for the determination of specific compounds in air by thermal desorption/gas chromatography [1-3]. One of the primary differences in these methods is the sorbents used in the thermal desorption tubes.					

REAGENTS:

1. Air, dry
2. Helium, high purity
3. Organic compounds of interest for mass spectra verification (See Table 1).*
4. Solvents for preparing spiking solutions: carbon disulfide (low benzene chromatographic grade), methanol, etc.(99+% purity)

* See SPECIAL PRECAUTIONS

EQUIPMENT:

1. Sampler: Thermal sampling tube, ¼" s.s. tube, multi-bed sorbents capable of trapping organic compounds in the C₃-C₁₆ range. Exact sampler configuration depends on thermal desorber system used. See Figure 1 for example.
2. Personal sampling pump, 0.01 to 0.05 L/min, with flexible tubing.
3. Shipping containers for thermal desorber sampling tubes.
4. Instrumentation: thermal desorption system, focusing capability, desorption temperature appropriate to sorbents in tube (~300 °C), and interfaced directly to a GC-MS system.
5. Gas chromatograph with injector fitted with 1/4" column adapter, 1/4" Swagelok nuts and Teflon ferrules (or equivalent).
6. Syringes: 1-µL, 10-µL (liquid); 100-µL, 500-µL (gas tight)
7. Volumetric Flasks, 10-mL.
8. Gas bulb, 2 L

SPECIAL PRECAUTIONS: Some solvents are flammable and should be handled with caution in a fume hood. Precautions should be taken to avoid inhalation of the vapors from solvents as well. Skin contact should be avoided.

SAMPLING:

NOTE: Prior to field use, clean all thermal desorption tubes thoroughly by heating at or above the intended tube desorption temperature for 1-2 hours with carrier gas flowing at a rate of at least 50 mL/min. Always store tubes with long-term storage caps attached, or in containers that prevent contamination. Identify each tube uniquely with a permanent number on either the tube or tube container. Under no circumstances should tape or labels be applied directly to the thermal desorption tubes.

1. Calibrate each personal sampling pump with a representative sampler in line.
2. Remove the caps of the sampler immediately before sampling. Attach sampler to personal sampling pump with flexible tubing.

NOTE: With a multi-bed sorbent tube, it is extremely important to sample in the correct direction, from least to maximum strength sorbent.

3. For general screening, sample at 0.01 to 0.05 L/min for a maximum sample volume of 6 L. Replace caps immediately after sampling. Keep field blanks capped at all times. Tubes can act as diffusive samplers if left uncapped in a contaminated environment.
4. Collect a "humidity test" sample to determine if the thermal adsorption tubes have a high water background.

NOTE: At higher sample volumes, additional analyte and water (from humidity) may be collected on the sampling tube. At sufficiently high levels of analyte or water in the sample, the mass spectrometer may malfunction during analysis resulting in loss of data for a given sample.

5. Collect a "control" sample. For indoor air samples this could be either an outside sample at the same location or an indoor sample taken in a non-complaint area.
6. Ship in sample storage containers at ambient temperature. Store at -10 °C.

SAMPLE PREPARATION:

7. Allow samples to equilibrate to room temperature prior to analysis. Remove each sampler from its storage container.

8. Analyze "humidity test" sampler first to determine if humidity was high during sampling (step 10).
9. If high humidity, dry purge the tubes with purified helium at 50 to 100 mL/min for a maximum of 3 L at ambient temperature prior to analysis. .
10. Place the sampler into the thermal desorber. Desorb in reverse direction to sampling flow.

CALIBRATION AND QUALITY CONTROL:

11. Tune the mass spectrometer according to manufacturer's directions to calibrate.
12. Make at least one blank run prior to analyzing any field samples to ensure that the TD-GC-MS system produces a clean chromatographic background. Also make a blank run after analysis of heavily concentrated samples to prevent any carryover in the system. If carryover is observed, make additional blank runs until the contamination is flushed from the thermal desorber system.
13. Maintain a log of thermal desorber tube use to record the number of times used and compounds found. If unexpected analytes are found in samples, the log can be checked to verify if the tube may have been exposed to these analytes during a previous sampling use.
14. Run spiked samples along with the screening samples to confirm the compounds of interest. To prepare spiked samples, use the procedure outlined in the Appendix .

MEASUREMENT:

15. See Appendix for conditions. MS scan range should cover the ions of interest, typically from 20 to 300 atomic mass units (amu). Mass spectra can either be identified by library searching or by manual interpretation (see Table 1). In all cases, library matches should also be checked for accurate identification and verified with standard spikes if necessary.

EVALUATION OF METHOD:

The method has been used for a number of field screening evaluations to detect volatile organic compounds. Estimate of the limit of detection for the method is based on the analysis of spiked samples for a number of different types of organic compounds. For the compounds studied, reliable mass spectra were collected at a level of 100 ng per compound or less. In situations where high levels of humidity may be present on the sample, some of the polar volatile compounds may not be efficiently collected on the internal trap of the thermal desorber. In these situations, purging of the samples with 3 L of helium at 100 mL/min removed the excess water and did not appreciably affect the recovery of the analytes on the sample.

REFERENCES:

- [1] Health and Safety Executive [1992]. MDHS 72 - Volatile organic compounds in air. Methods for the determination of hazardous substances. HMSO: London: ISBN 0-11-885692-8.
- [2] McCaffrey CA, MacLachlan J, Brookes BI [1994]. Adsorbent tube evaluation for the preconcentration of volatile organic compounds in air for analysis by gas chromatography-mass spectrometry. Analyst 119:897-902.
- [3] Bianchi AP, Varney MS [1992]. Sampling and analysis of volatile organic compounds in estuarine air by gas chromatography and mass spectrometry. J. Chromatogr. 643:11-23.
- [4] EPA [1984]. Environmental Protection Agency Air Toxics Method T01. Rev. 1.0 (April, 1984): Method for the determination of volatile organic compounds in ambient air using Tenax(R) adsorption and gas chromatography/mass spectrometry (GC/MS), Section 13.

METHOD WRITTEN BY:

Ardith A. Grote and Eugene R. Kennedy, Ph.D., NIOSH, DPSE

TABLE 1. COMMON VOLATILE ORGANIC COMPOUNDS WITH MASS SPECTRAL DATA

Compound /Synonyms	CAS# RTECS	Empirical Formula	MW ^a	BP ^b (°C)	VP ^c @ 25 °C mm Hg kPa		Characteristic Ions, m/z
Aromatic Hydrocarbons							
Benzene /benzol	71-43-2 CY1400000	C ₆ H ₆	78.11	80.1	95.2	12.7	78*
Xylene /dimethyl benzene	1330-20-7 ZE2100000	C ₈ H ₁₀	106.7				91, 106*, 105
o-xylene				144.4	6.7	0.9	
m-xylene				139.1	8.4	1.1	
p-xylene				138.4	8.8	1.2	
Toluene /toluol	108-88-3 XS5250000	C ₇ H ₈	92.14	110.6	28.4	3.8	91, 92*
Aliphatic Hydrocarbons							
n-Pentane	109-66-0 RZ9450000	C ₅ H ₁₂	72.15	36.1	512.5	68.3	43, 72*, 57
n-Hexane /hexyl-hydride	110-54-3 MN9275000	C ₆ H ₁₄	86.18	68.7	151.3	20.2	57, 43, 86*, 41
n-Heptane	142-82-5 MI7700000	C ₇ H ₁₆	100.21	98.4	45.8	6.1	43, 71, 57, 100*, 41
n-Octane	111-65-9 RG8400000	C ₈ H ₁₈	114.23	125.7	14.0	1.9	43, 85, 114*, 57
n-Decane /decyl hydride	124-18-5 HD6500000	C ₁₀ H ₂₂	142.29	174	1.4	0.2	43, 57, 71, 41, 142*
Ketones							
Acetone /2-propanone	67-64-1 AL3150000	C ₃ H ₆ O	58.08	56	266	35.5	43, 58*
2-Butanone /methyl ethyl ketone	78-93-3 EL6475000	C ₄ H ₈ O	72.11	79.6	100	13	43, 72*
Methyl isobutyl ketone /MIBK, hexone	108-10-1 SA9275000	C ₆ H ₁₂ O	100.16	117	15	2	43, 100*, 58
Cyclohexanone /cyclohexyl ketone	108-94-1 GW1050000	C ₆ H ₁₀ O	98.15	155	2	0.3	55, 42, 98*, 69
Alcohols							
Methanol /methyl alcohol	67-56-1 PC1400000	CH ₃ OH	32.04	64.5	115	15.3	31, 29, 32*
Ethanol /ethyl alcohol	64-17-5 KQ6300000	C ₂ H ₅ OH	46.07	78.5	42	5.6	31, 45, 46*
Isopropanol /1-methyl ethanol	67-63-0 NT8050000	C ₃ H ₇ OH	60.09	82.5	33	4.4	45, 59, 43
Butanol /butyl alcohol	71-36-3 EO1400000	C ₄ H ₉ OH	74.12	117	4.2	0.56	56, 31, 41, 43

Compound /Synonyms	CAS# RTECS	Empirical Formula	MW ^a	BP ^b (°C)	VP ^c @ 25 °C mm Hg	kPa	Characteristic Ions, m/z
Glycol Ethers							
Butyl cellosolve /2-butoxyethanol	111-76-2 KJ8575000	C ₈ H ₁₄ O ₂	118.17	171	0.8	0.11	57, 41, 45, 75, 87
Diethylene glycol ethyl ether /Carbitol	111-90-0 KK8750000	C ₆ H ₁₄ O ₃	134.17	202	0.08	0.01	45, 59, 72, 73, 75, 104
Phenolics							
Phenol /hydroxybenzene	108-95-2 SJ3325000	C ₆ H ₅ OH	94.11	182	47	0.35	94*, 65, 66, 39
Cresol	1319-77-3 GO5950000	C ₇ H ₇ OH	108.14				108*, 107, 77, 79
2-methylphenol	95-48-7			190.9	1.9	0.25	
3-methylphenol	108-39-4			202.2	1.0	0.15	
4-methylphenol	106-44-5			201.9	0.8	0.11	
Chlorinated Hydrocarbons							
Methylene chloride /dichloromethane	75-09-2 PA8050000	CH ₂ Cl ₂	84.94	40	349	47	86*, 84, 49, 51
1,1,1-Trichloroethane /methyl chloroform	71-55-6 KJ2975000	CCl ₃ CH ₃	133.42	75	100	13.5	97, 99, 117, 119
Perchloroethylene /hexachloroethane	127-18-4 KX3850000	CCl ₃ CCl ₃	236.74	187 (subl)	0.2	<0.1	164*, 166, 168, 129, 131, 133, 94, 96
o-,p- Dichlorobenzenes		C ₆ H ₄ Cl ₂	147.0				146*, 148, 111, 113, 75
/1,2-dichlorobenzene	95-50-1 CZ4500000			172-9	1.2	0.2	
/1,4- dichlorobenzene	106-46-7 CZ4550000			173.7	1.7	0.2	
1,1,2-Trichloro-1,2,2- trifluoroethane /Freon 113	76-13-1 KJ4000000	CCl ₂ FCFClF ₂	187.38	47.6	384	38	101, 103, 151, 153, 85, 87
Terpenes							
d-Limonene	5989-27-5 OS8100000	C ₁₀ H ₁₆	136.23	176	1.2		68, 67, 93, 121, 136*
Turpentine (Pinenes)	8006-64-2	C ₁₀ H ₁₆	136.23	156 to 170	4 @ 20°		93, 121, 136*, 91
α-pinene	80-56-8			156			
β-pinene	127-91-3			165			
Aldehydes							
Hexanal /caproaldehyde	66-25-1 MN7175000	C ₆ H ₁₂ O	100.16	131	10	1.3	44, 56, 72, 82, 41

Compound /Synonyms	CAS# RTECS	Empirical Formula	MW ^a	BP ^b (°C)	VP ^c @ 25 °C mm Hg	kPa	Characteristic Ions, m/z
Benzaldehyde /benzoic aldehyde	100-52-7 CU4375000	C ₇ H ₁₂ O	106.12	179	1.0	0.1	77, 105, 106*, 51
Nonanal /pelargonic aldehyde	124-19-6 RA5700000	C ₉ H ₁₈ O	142.24	93	23	3	43, 44, 57, 98, 114
Acetates							
Ethyl acetate /acetic ether	141-78-6 AH5425000	C ₄ H ₈ O ₂	88.1	77	73	9.7	43, 88*, 61, 70, 73, 45
Butyl acetate /acetic acid butyl ester	123-86-4 AF7350000	C ₈ H ₁₂ O ₂	116.16	126	10	1.3	43, 56, 73, 61
Amyl acetate /banana oil	628-63-7 AJ1925000	C ₇ H ₁₄ O ₂	130.18	149	4	0.5	43, 70, 55, 61
Other							
Octamethylcyclotetra- siloxane	556-67-2 GZ4397000	C ₈ H ₂₄ O ₄ Si ₄	296.62	175			281, 282, 283

^a Molecular Weight

^b Boiling Point

^c Vapor Pressure

* Indicates molecular ion

APPENDIX

Multi-bed sorbent tubes: Other sorbent combinations and instrumentation/conditions shown to be equivalent may be substituted for those listed below. In particular, if the compounds of interest are known, specific sorbents and conditions can be chosen that work best for that particular compound(s). The tubes that have been used in NIOSH studies with the Perkin Elmer ATD system are ¼" stainless steel tubes, and are shown in the diagram below:




Figure 1

Carbopack™ and Carboxen™ adsorbents are available from Supelco, Inc.

Preparation of spiked samples Spiked tubes can be prepared from either liquid or gas bulb standards.

Liquid standards: Prepare stock solutions by adding known amounts of analytes to 10-mL volumetric flasks containing high purity solvent (carbon disulfide, methanol, toluene). Solvents are chosen based on solubility for the analytes of interest and ability to be separated from the analytes when chromatographed. Highly volatile compounds should be dissolved in a less volatile solvent. For most compounds, carbon disulfide is a good general purpose solvent, although this will interfere with early eluting compounds.

Gas bulb standards: Inject known amounts of organic analytes of interest into a gas bulb of known volume filled with clean air [4]. Prior to closing the bulb, place a magnetic stirrer and several glass beads are placed in the bulb to assist in agitation after introduction of the analytes. After injection of all of the analytes of interest into the bulb, warm the bulb to 50 °C and place it on a magnetic stirring plate and stir for several minutes to ensure complete vaporization of the analytes. After the bulb has been stirred and cooled to room temperature, remove aliquots from the bulb with a gas syringe and inject into a sample tube as described below.

Tube spiking Fit a GC injector with a ¼" column adapter. Maintain the injector at 120 °C to assist in vaporization of the injected sample. Attach cleaned thermal desorption tubes to injector with ¼" Swagelok nuts and Teflon ferrules, and adjust helium flow through the injector to 50 mL/min. Attach the sampling tube so that flow direction is the same as for sampling. Take an aliquot of standard solution (gas standards 100 to 500 µL; liquid standards, 0.1 to 2 µL) and inject into the GC injector. Allow to equilibrate for 10 minutes. Remove tube and analyze by thermal desorption using the same conditions as for field samples.

Instrumentation: Actual media, instrumentation, and conditions used for general screening of unknown environments are as follows: Perkin-Elmer ATD 400 (automated thermal desorption system) interfaced directly to a Hewlett-Packard 5980 gas chromatograph/HP5970 mass selective detector and data system.

ATD conditions:

Tube desorption temperature: 300°C
Tube desorption time: 10 min.
Valve/transfer line temperatures: 150°C
Focusing trap: Carbopack B/Carboxen 1000, 60/80 mesh, held at 27°C during tube desorption
Focusing trap desorption temperature: 300°C
Desorption flow: 50-60 mL/min.
Inlet split: off
Outlet split: 20 mL/min.
Helium: 10 PSI

GC conditions:

DB-1 fused silica capillary column, 30 meter, 1- μ m film thickness, 0.25-mm I.D.
Temperature program: Initial 35°C for 4 minutes, ramp to 100°C at 8°/min., then ramp to 300°C at 15°/min, hold 1-5 minutes.
Run time: 27 min.

MSD conditions:

Transfer line: 280°C
Scan 20-300 amu, EI mode
EMV: set at tuning value
Solvent delay: 0 min. for field samples; if a solvent-spiked tube is analyzed, a solvent delay may be necessary to prevent MS shutdown caused by excessive pressure.